

FORMAT B.1					
Format for Technical Pre-Bid Queries					
Tender No		TPSODL/OT/2021-22/037			
Package Name		Rate Contract of Meter Test Bench			
Sr. No.	Page No	Detailed Reference to TPSODL Technical Document. Please specify Document No / Clause No / Page No	Description as per Bid Document	TPSODL Response	
1			Vide Pre-bid queries Sr. No. 19, Burden Measurement Function – We understand that Burden measurement function is implied need of the testing as per Standard IS13779/IS14697/IS16444. Hence it is required to provide such facility with the asked test bench using necessary hardware and software for burden measurement. Hence we understand that specifications mentioned under Sr. No. 19 of Pre-bid Queries Reply will be part of the asked technical specifications to have this functionality and also it will be verify during technical evaluation from all the bidders.	Burden measurement function is implied need of the testing, hence it is required to provide such facility with the asked meter test bench using necessary hardware and software for burden measurement. It is part of the asked technical specifications and such tests will be verified during evaluation of all the bidders	
2			Vide Clause No. 4.3 Test to be Performed as per IS15959 Part 1, 2, 3 & IS16444 Part 1, 2 – we understand that the asked test bench is extensively use to check the functionality of Smart meters including Metrology & Communication, hence it is the implied need to include the functional testing as per mentioned in the standard IS15959/IS16444 for smart meter functionality. Also we understand that the communication & logging function of Smart meter such temper events, power events etc. shall be also check in the presence of possible field influence such as low voltages, harmonics etc. We are enclosing the detail specifications as per Annexure-A along with the specifications floated by various Indian utilities in regards of SMART meters functional testing for your reference. Also these features will be verify during technical evaluation from all the bidders.	Smart meters functional testing as per technical specifications is required and the said functionality will also be verified during evaluation of all the bidders	
3			After warranty period AMC for 5 years has been asked. Please confirm Non Comprehensive AMC required or Comprehensive AMC Required.	Comprehensive AMC	
4			Make in India (MII) procurement policy issued from the government is applicable for this tender. Since it supports our nation for become self-dependent. Please confirm.	As per tender document Qualification Criteria	